

IMPORTANT NOTE

BID DOCUMENT HAS BEEN DISPLAYED BELOW TO UNDERSTAND THE REQUIREMENT ONLY. PARTIES INTERESTED TO PARTICIPATE AGAINST SUCH REQUIREMENTS MAY APPROACH WITH COMPLETE CREDENTIALS TO THE FOLLOWING OFFICE:

HEAD (CALCUTTA BRANCH)
OIL INDIA LIMITED
4, INDIA EXCHANGE PLACE
KOLKATA – 700 001
E-MAIL: oilcalmn@cal2.vsnl.net.in
FAX: 033-22302596

ISSUE OF ENQUIRY TO THE PARTY MAY BE CONSIDERED AGAINST PRESENT/ FUTURE TENDER BASED ON THE CREDENTIALS.

OIL INDIA LIMITED

(A Govt. Of India Enterprise) Tel :033 2230 1657, 1658
 4, India Exchange Place, Fax :91 33 2230 2596
 Kolkata-700001 E-mail :oilcalmn@cal2.vsnl.net.in

Tender No. & Date : KID1975L09/09 29.01.2009

Bid Security Amount : INR 0.00 OR USD 0.00
 (or equivalent Amount in any currency)

Bidding Type : Single Bid (Composite Bid)

Bid Closing On : 06.03.2009 at 14:00 hrs. (IST)
 Bid Opening On : 06.03.2009 at 14:00 hrs. (IST)

Performance Guarantee : Not Applicable

OIL INDIA LIMITED invites Limited tenders for items detailed below:

Item No./ Mat. Code	Material Description	Quantity	UOM
10 0C000172	<p>BLOOD AND BODY FLUID CULTURE SYSTEM TECHNICAL SPECIFICATION</p> <ol style="list-style-type: none"> 1. IT SHOULD BE FULLY AUTOMATED WALK-AWAY & CONTINUOUS MONITORING ONE. 2. SYSTEM SHOULD HAVE CONTINUOUS AGITATION FOR OPTIMIZED RECOVERY OF ORGANISMS. 3. DETECTION PRINCIPLE SHOULD BE BASED ON SENSITIVE FLOURESCENT TECHNOLOGY. 4. SYSTEM SHOULD HAVE NON-INVASIVE TECHNOLOGY i.c. NON PUNCTURING OF THE SAMPLE BOTTLES DURING SAMPLE INTERPRETATION. 5. SYSTEM SHOULD HAVE CONTINUOUIS AGITATION, TO ENSURE FASTER GROWTH OF ORGANISMS. 6. SYSTEM SHOULD MONITOR EACH SAMPLE POSITIONS IN EVERY 10 MINUTES. 7. SYSTEM SHOULD HAVE PROVISION OF INCUBATION TEMPERATURE FROM 35-37 DIGREE CENTRIGRADE. 8. SYSTEM SHOULD HAVE CAPABILITY OF ANALYZING DELAYED VIAL ENTRY UP TO 48 HOURS STORED AT ROOM TEMPERATURE. 9. SYSTEM SHOULD HAVE MINIMUM 50 SAMPLE POSITIONS. 10. SYSTEM SHOULD HAVE SIMPLE HANDLING PROCEDURE. 11. SYSTEM SHOULD HAVE AUTO QC FACILITY TO AVOID ANY MANUAL DAILY MAINTENANCE. 12. SYSTEM SHOULD HAVE MORE THAN 12 MEDICA SPECIFIC ALGORITHMS TO ENSURE EARLY DETECTION OF MICRO-ORGANISMS. 13. SHOULD HAVE SPECIAL MEDIA FOR PEDIATRIC SAMPLES AND STERILE BODY FLUID SAMPLES. 14. SHOULD HAVE SPECIAL LYTIC ANAEROBIC MEDIA FOR INCREASED DETECTION OF PARTIALLY PHOGOCYTISED-ORGANISMS. 15. SHOULD HAVE SPECIAL MEDIA FOR OPTIMAL RECOVERY OF YEAST, FUNGI AND MYCOBACTERIUM FROM BLOOD SAMPLES. 16. SHOULD HAVE SPECIAL SUPPLEMENT FOR ENHANCED RECOVERY OF LOW VOLUME STERILE BODY FLUID. 	1	NO

Tender No. & Date : KID1975L09/09 29.01.2009

Item No./ Mat. Code	Material Description	Quantity	UOM
	17. SHOULD HAVE SPECIAL ANTIBIOTIC NEUTRALIZATION RESN MEDIA. 18. ANTIBIOTIC NEUTRALIZING MEDIA SHOULD NOT HAVE ANY MASKINGEFFECT IN GRAM-STAINING OF POSITIVE SAMPLES FOR EASE OF GRAM STAINING INTERPRETATION TO RELEASE PRIMARY REPORT(S). 19. SYSTEM SHOULD BE SUPPLIED ALONG WITH 1 KVA UPS WITH 30 MINUTES BACK-UP.		
<u>20</u> 0C000172	TUBE SEALER. SPECIFICATION. 1. RADIO FREQUENCY SEALING - THE RADIO FREQUENCY WAVE GENERATED BY AN OPTIMALLY DESIGNED RF OSCILLATOR GUARANTEES EXCEPTIONAL SEAL QUALITY. 2. DESIGNED TO MEET ALL INTERNATIONAL SAFETY REQUIREMENTS OF EN61010- 1. ENSURES SAFETY AGAINST ELECTRICAL SHOCK HAZARDS, FIRE HAZARD MECHANICAL HAZARDS, ELECTRO MAGNETIC INTERFERENCE ETC. 3. NO HAEMOLYSIS OF BLOOD IN THE TUBE SEGMENTS AS RADIO FREQUENCY SEALING HEATS ONLY THE TUBE AND NOT BLOOD INSIDE. 4. EASY SEPARATION OF TUBE SEGMENTS AFTER SEALING. 5. PRODUCES A HERMETIC SEAL HENCE NO CONTAMINATION. 6. EQUIPPED WITH SMPS TO WITHSTAND WIDE INPUT VOLT. FLUCTUATIONS. SMPS ENSURES THAT THE OUTPUT VOL. REMAINS CONSTANT AT A SET VOLT. IRRESPECTIVE OF THE INPUT VOLT. HENCE, THE SEAL QUALITY REMAINS THE SAME IRRESPECTIVE OF VOLT. FLUCTUATIONS. 7. EASY TO CLEAN AS THE ELECTRODES ARE EASILY ACCESSIBLE. 8. CONFORMS TO STANDARDS AND SPECIFICATIONS OF ISO 9001:2000, CE MARKING AND CFSA MARKING.	1	NO

Special Notes : 1. The bidder must submit technical literature / catalogue of the offered product in duplicate along with the offer, failing which offer may be liable for rejection.
 2. Guarantee / Warranty certificate for one year will be required along with the supply.
 3. Packing must be adequate to avoid transit damages.

Tender No. : KID1975L09/09
Tender Date : 29.01.2009
Bid Closing On : 06.03.2009 at 14:00 hrs.(IST)
Bid Opening On : 06.03.2009 at 14:00 hrs.(IST)

Tender issued to following parties only:

Slno	V_Code	Vendor Name	City/Country
1	200246	SURGICAL MART	DIBRUGARH
2	200795	HOSPITAL SUPPLY CO	KOLKATA
3	202450	HOSPIMEDICA INTERNATIONAL LTD.	NEW DELHI
4	203253	CHIMCO BIO-MEDICAL ENGINEERING COMP	MUMBAI
5	203731	MEDITRIN INSTRUMENTS	GUWAHATI
6	203813	VISHAL SURGICAL EQUIPMENT CO PVT LT	CHENNAI
7	203943	SURGIQUIP SYSTEMS,	HARYANA.
8	204234	TERUMO PENPOL	KOLKATA
9	204235	TERUMO PENPOL	GUWAHATI
10	204522	MEDISPHERE MARKETING LTD	NEW DELHI
11	205193	BIOMERIEUX INDIA PVT. LTD.	KOLKATA