



OIL INDIA LIMITED
(A Govt. of India Enterprise)
P.O. Duliajan-786602, Assam
E-mail:material@oilindia.in, **Fax No.**91-374-2800533

OIL INDIA LIMITED invites sealed tenders for the following:

Tender No.	B. C. Date	Material Description
DID2034P12/NP	14.07.2011	ORTHPANTOMOGRAPHY WITH CEPHALOMETRY = 1 No.

2.0 Bid documents (Non transferable) can be purchased from 14.06.2011 till one day prior to the respective B.C. Dates on payment of tender fee of Rs.1,000.00 (excepting for PSUs and SSI units registered for the item) each through Crossed Demand Draft in favour of M/s. Oil India Limited payable at the place of purchase of tender document from

- Head - Materials, Oil India Limited, P.O. Duliajan, Assam -786602
- Head -Calcutta Branch, Oil India Limited, 4,India Exchange Place, Kolkata - 700001
- Senior Adviser(Contract & Purchase), Oil India Limited, Plot No. 19, Sector- 16 A, NOIDA-201301
- Addl. Chief Materials Manager (Pipeline), Oil India Limited,PO: Udayan Vihar, Guwahati - 781171.

3.0 Detailed tender document will be available in OIL's website
www.oil-india.com

4.0 To be eligible for issue of tender documents, the applicant must meet the following qualifying criteria (documentary evidence to be provided) :

i. Successful execution of a single order of value not less than the amount shown below for supply of similar items during last five years:

Tender No.	Single Order Value in Lakhs (Rs.)
DID2034P12/NP	15.00

ii. Annual turnover of the firm in any of the last three financial years or current financial year should be more than the amount as below :

Tender No.	Annual Turnover in Lakhs (Rs.)
DID2034P12/NP	30.00

OIL INDIA LIMITED(A Govt. of India Enterprise)
P.O. Duliajan-786602, Assam**E-mail:material@oilindia.in, Fax No.91-374-2800533****Tender No. & Date : DID2034P12/NP 18.05.2011**Tender Fee : INR 1,000.00
Bid Security Amount : INR 60,000.00**Bidding Type : Single Bid (Composite Bid)**Bid Closing On : 14.07.2011 at 13:00 hrs. (IST)
Bid Opening On : 14.07.2011 at 13:00 hrs. (IST)

Performance Guarantee : Applicable

OIL INDIA LIMITED invites Press tenders for items detailed below:

Item No./ Mat. Code	Material Description	Quantity	UOM
10 0C000172	<p>OTHOPANTOMOGRAPHY WITH CEPHALOMETRY :</p> <p>SPECIFICATION OF X-RAY TUBE :</p> <ul style="list-style-type: none"> - GENERATOR : 130 K Hz OR ABOVE - LINE VOLTAGE : 230 V/240 V 50 Hz - TUBE VOLTAGE : 60 - 90 kV (1 kV STEP) - TUBE CURRENT : 2 - 15mA - EXPOSURE COMPENSATION : AUTOMATIC EXPOSURE CONTROL - MUST BE A DIGITAL PROCESS. - FOCAL SPOT SIZE : 0.5 MM AS PER IEC STANDARDS - TOTAL FILTRATION : EQUIVALENT TO 2.5 MM AI OR MORE <p>SENSORS :</p> <ul style="list-style-type: none"> - HIGH RESOLUTIONH CCD SENSORS WITH FIBER OPTIC LAYER FOR PROTECTION AND LONGEVITY OF THE SENSOR. <p>IMAGING PROCEDURE :</p> <ul style="list-style-type: none"> - USING USER FRIENDLY GRAPHICAL USER INTERFACE ON A COMPUTER AND FROM THE CONSOLE ON THE MACHINE ITSELF. - THE SOFTWARE MUST BE USER FRIENDLY, SHOULD BE ABLE TO VOCALLY ARTICULATE THE CHOICES DURING THE USE, ALL THE PARAMETERS LIDE kV mA SHOULD BE SELECTABLE, IRRESPECTIVE OF EACH OTHER. <p>PROGRAMME : PANOROMIC (ADULT AND CHILD), SEGMENTED PANOROMIC, LATERAL TMJ, (2 & 4 VIEWS), MAXILLARY SINUS.</p> <ul style="list-style-type: none"> - MAXIMUM TIME FROM THE SCANNING TO IMAGE SHOULD NOT BE MORE THAN 15 SEC FOR ADULT AND 13.5 SEC FOR CHILD TO ENSURE LOW DOSAGE. PATIENT POSITIONING : THE POSITIONING MUST BE OPEN WHERE THE DENTIST SHOULD BE ABLE TO FACE THE PATIENT FOR BETTER PATIENT COMFORT AND ACCURATE POSITIONING. - THE PRODUCT MUST BE COMPACT AND SHOULD BE ABLE TO BE FITTED IN A MAXIMUM SPACE OF 4.5 X 4.5 FT. MAXIMUM. - THE SOFTWARE MUST BE READY FOR DIRECT ACQUISITION OF 	1	NO

Tender No. & Date : DID2034P12/NP

18.05.2011

Item No./ Mat. Code	Material Description	Quantity	UOM
	<p>IMAGES FROM RVG, INTRA ORAL CAMERA, EXTRA ORAL CAMERA, CEPHALOMETRY, AND DIGITAL PHOTOGRAPHY.</p> <ul style="list-style-type: none"> - THE PATIENT INFORMATION MUST BE ARRANGED IN A VERY USER FRIENDLY AND SIMPLE FORMAT WITH BUILT IN PATIENT INFORMATION MANAGEMENT SERVICE. - THE IMAGES SHOULD BE ABLE TO BE STORED IN DICOM FORMAT FOR UNIVERSAL USE. <p>CHEPH UNIT :</p> <ul style="list-style-type: none"> - THE UNIT MUST BE LOADED WITH ITS OWN DIGITAL DETECTOR, NEED NOT CHANGE THE DIGITAL SENSOR POSITION FROM OPG TO CEPH. - THE EXPOSURE TO THE PATIENT SHOULD NOT BE MORE THAN 2 SEC. - THE POSITIONING OF PATIENT SHOULD BE SIMPLE, AND ALL THE POSITIONS SHOULD BE POSSIBLE TO BE PROGRAMMED IN THE GRAPHICAL USER INTERFACE INCLUDING ANGULAR CEPHALOMETRY POSITIONS. - THE PATIENT INFORMATION MUST BE ARRANGED IN A VERY USER FRIENDLY AND SIMPLE FORMAT WITH BUILT IN PATIENT INFORMATION MANAGEMENT SERVICE. - THE IMAGES SHOULD BE ABLE TO BE STORED IN DICOM FORMAT FOR UNIVERSAL USE. - THE CEPHALOMETRY ARM SHOULD NOT BE MORE THAN 3 METERS IN LENGTH. <p>LASER MEDICAL CAMERA :</p> <ul style="list-style-type: none"> - MEDICAL GRADE LASER CAMERA. - RESOLUTION 550 dpi OR ABOVE. - SPEED OF MINIMUM 150 FOR 14 X 17 FILM SIZE - TOUCH SCREEN FOR EASY OPERATION. - FLAXIBILITY OF USING DIFFERENT SIZES OF FILMS IN ANY OF THE DRAWERS. - MINIMUM 2 DRAWERS. - FILMS SHOULD BE LOADED IN DAYLIGHT, AND HALF USED CARTRIDGE CAN BE REMOVED, IF NEEDED. - POWER 230 VAC 50 Hz. - SHOULD BE FULLY COMPATIBLE WITH ANY DICOM MODALITY viz. DENTAL CT, DIGITAL OPG, CEPH, CR ETC. 		

Standard Notes: (1) VALIDITY : Your offer must be valid for 120 days from the date of bid opening. Offer with inadequate validity will be rejected.

(2) The offer should be submitted in triplicate.

(3) In the event of receipt of only a single offer against the tender within B.C. date, OIL reserves the right to extend the B.C. date as deemed fit by the Company. During the extended period, the bidders who have already submitted the bids on or before the original B.C. date, shall not be permitted to revise their quotation.

(4) Any sum of money due and payable to the contractor (including Security Deposit refundable to them) under this or any other contract may be appropriated by Oil India

Tender No. & Date : DID2034P12/NP 18.05.2011

Limited and set-off against any claim of Oil India Limited (or such other person or persons contracting through Oil India Limited) for payment of a sum of money arising out of this contract or under any other contract made by the contractor with Oil India Limited (or such other person or persons contracting through Oil India Limited).

(5) In the event you authorize your dealer/stockist/channel partner to quote on your behalf, the dealer/stockist/channel partner while submitting bid should mention on the body of the envelope that they are submitting bid on your behalf. In the event the dealer/stockist/channel partner do not mention the name of their OEM/principal on the body of the envelope, the offer shall be treated as unsolicited offer and will not be considered for opening. The dealer/stockist/channel partner should take note of above while submitting bid on behalf of their OEM/principal.

(6) For order with F.O.R. Destination term, 100% payment against despatch documents will not be entertained. In this regards please refer payment terms in ANNEXURE-MM/TENDER/LP/01/06.

(7) To evaluate the inter-se ranking of the offers, Assam Entry Tax on purchase value will be loaded as per prevailing Government of Assam Guidelines as applicable on bid closing date. Bidders may check this with the appropriate authority while submitting their bids.