

IMPORTANT NOTE

BID DOCUMENT HAS BEEN DISPLAYED BELOW TO UNDERSTAND THE REQUIREMENT ONLY. PARTIES INTERESTED TO PARTICIPATE AGAINST THIS TENDER SHALL HAVE TO PURCHASE THE TENDER DOCUMENT FROM ANY OF OIL'S DESIGNATED OFFICES MENTIONED IN THE TENDER NOTIFICATION GIVEN BELOW. PROOF OF PURCHASE OF TENDER DOCUMENT MUST BE SUBMITTED ALONG WITH THE OFFER FAILING WHICH OFFERS WILL BE TREATED AS UNSOLICITED.

OIL INDIA LIMITED

(A Govt. Of India Enterprise) Tel :033 2230 1657, 1658
 4, India Exchange Place, Fax :91 33 2230 2596
 Kolkata-700001 E-mail :oilcalmn@cal2.vsnl.net.in

Tender No. & Date : KID7096P08/09 06.11.2007

Tender Fee : INR 1,000.00
 Bid Security Amount : INR 60,000.00

Bidding Type : Single Bid (Composite Bid)

Bid Closing On : 08.01.2008 at 13:00 hrs. (IST)
 Bid Opening On : 08.01.2008 at 13:00 hrs. (IST)

Performance Guarantee : Applicable

OIL INDIA LIMITED invites Press tenders for items detailed below:

Item No./ Mat. Code	Material Description	Quantity	UOM
10 0C000677	<p>COMPUTERISED RADIOGRAPHY SYSTEM FOR OIL INDIA LIMITED'S HOSPITAL AT DULIAJAN, ASSAM:</p> <p>TECHNICAL SPECIFICATIONS</p> <p>1] DIRECT DIGITIZER</p> <p>Exposure Size: 14# x 17#, 14# x 14#, 14#x 11#, 12#x 10#, 10#x 8#, 18 x 24 cms and others Sampling Pitch: 3 types: 87.5 , 175 and 43.75 µm for Mammography Maximum Resolution:4020 x 4892, 5440 x 6776, 7080 x 9480 Digital Gradation Levels: 4096 levels(12 bit) Processing capability: Approx. 90 plates/hour Cassette feed / Load time or Cassette Cycle Time : Approx. 40 secs. Slots: At least, Insert x 1(plus 1 stack) Eject x 1 (4 stack) Power Consumption: AC 100/115/120/200/220/230/240 V ± 10% 50/60 Hz approx. 1.1 kW Operating conditions : Temp.: 15 # 300 C , Humidity : 40-80% RH (No condensation) Console: Image Processing: Automatic gradation processing, Frequency processing, Equalization processing, Hybrid processing Image Output: Host: 3 ch for normal, 1 ch for backup. Printer: 2 ch for normal, 1 ch for backup Standard software: Host or Printer: 1 ch for normal, 1 ch for backup. DICOM Support: Basic Grayscale print management, Storage, Modality worklist management, Modality performed procedure step, Grayscale standard display function Output format: 1 x 1, 2 x 1, 1 x 2 Control Console: PC with Medical Grade High resolution Touch Screen LCD Display Software Option: Mammography option, Stitching option Cassettes to be offered: At least 2 Nos. 14# x 17#, 2 Nos. 10# x 8# and 2 Nos.12# x 10#</p> <p>2] LASER CAMERA:</p>	1	NO

Tender No. & Date : KID7096P08/09

06.11.2007

Item No./ Mat. Code	Material Description	Quantity	UOM
	<p>Source: Semiconductor Laser Film Size: 14# x 17#, 14# x 14#, 11# x 14#, 10# x 12# and 8# x 10# selectable Film : Dry image recorded film Image format: 1,2,4,6,8,9,12,15,16,20,24,25,30,35,36,42,48,54,56,60,63,64 Image memory: Hard Disk 80 GB standard, Print memory 512 MB standard Input Port: 16 ports (max) Number of pixels: 8079 x 9752 pixels for 14# x 17# Pixel Size: 43.75 μ (standard), 25μ (fine) optional 1 GB memory are required available for 14# x 11#, 10# x 12# and 8# x 10# films Image data input: 8 bits / 12 bits Output gradation: 16384 levels (14 bits) Image mode: Pixel replication / Function interpolation process Processing capability :Approx. 120 sheets (14# x 17#) / hour Protocol: DICOM Print Management Input interface: Ethernet 10 base-T/100 base-TX Supply: Max. 3 channels, 14 inch tray is the 1st tray, 2nd and 3rd trays are selectable(14# x 17#, 10# x 12#, 8# x 10#) Standby function: Start timer / Nighttime standby Border processing: Black / White Trimmed frame: To be Available Density Correction: Automatic via built in densitometer Positive/Negative: To be Available DICOM functions: Presentation LUT service class Requested Image size Noise level: Printing: 55dB or less, Stand By: 55 dB or less Operating conditions: 15-300C, 30-70% RH(no condensation) Power: 100-120 V AC/220-240 V AC, 9.0-7.5 A/4.5-4.0 A Heating generation: Approx. 1200 kJ/hour(approx. 286 kcal/hour) Foot Print: 0.43 m2 Accessories: Power Cable, Operation Manual, Cutter (for film loading) Options: Sorter, Cleaning sheet, 1 GB Print memory Print Sizes: 3 Online Trays: 14# x 17#, 11# x 14#, 8# x 10#</p> <p>PRICES: The quoted prices should be on F.O.R. destination basis.</p> <p>INSTALLATION: The installation should be done by successful bidder FREE OF COST at Oil India Ltd's Hospital at Duliagan, Assam.</p> <p>SERVICE: After sales service should be available in Dibrugarh / Guwahati</p> <p>ANNUAL MAINTENANCE CONTRACT: Prices for Annual Maintenance Contract must be quoted for 5 years after the Warranty period(including Spares and Cassettes).</p> <p>WARRANTY: Warranty should be for at least 12 months from the date of installation against any manufacturing defect or 18 months from the date of supply, whichever is earlier.</p> <p>DELIVERY: Should be within 60 days after receipt of confirmed order.</p>		

Special Notes : 1. The bidder must submit technical literature / catalogue of the offered product in duplicate along with the offer, failing which offer may be liable for rejection.