

IMPORTANT NOTE

BID DOCUMENT HAS BEEN DISPLAYED BELOW TO UNDERSTAND THE REQUIREMENT ONLY. PARTIES INTERESTED TO PARTICIPATE AGAINST THIS TENDER SHALL HAVE TO PURCHASE THE TENDER DOCUMENT FROM ANY OF OIL'S DESIGNATED OFFICES MENTIONED IN THE TENDER NOTIFICATION GIVEN BELOW. PROOF OF PURCHASE OF TENDER DOCUMENT MUST BE SUBMITTED ALONG WITH THE OFFER FAILING WHICH OFFERS WILL BE TREATED AS UNSOLICITED.

OIL INDIA LIMITED

(A Govt. Of India Enterprise) Tel :033 2230 1657, 1658
 4, India Exchange Place, Fax :91 33 2230 2596
 Kolkata-700001 E-mail :oilcalmn@cal2.vsnl.net.in

Tender No. & Date : KID7439P08/09 05.12.2007

Tender Fee : INR 1,000.00
 Bid Security Amount : INR 0.00

Bidding Type : Single Bid (Composite Bid)

Bid Closing On : 29.01.2008 at 13:00 hrs. (IST)
 Bid Opening On : 29.01.2008 at 13:00 hrs. (IST)

Performance Guarantee : Not Applicable

OIL INDIA LIMITED invites Press tenders for items detailed below:

Item No./ Mat. Code	Material Description	Quantity	UOM
10 0C000172	<p>PHAECO EMMULSIFICATION MACHINE FOR USE IN OIL'S HOSPITAL DULIAJAN, ASSAM</p> <p>-----</p> <ol style="list-style-type: none"> 1. Fully programmable microprocessor based system. 2. Handpiece with minimum four piezoelectric crystals. 3. Should have cold phaco technology. 4. Handpiece should be compatible with tips like standard, microtip and Kelman. 5. Should have modality of continuous, short pulse, long pulse, high power pulse, low power pulse, burst mode (single/ multi/ burst continuous modes) 6. Anterior vitrectomy # electrical or guillotine cutter. 7. Should have preferably closed fluidic system or else cassette mechanism. 8. Should have flat screen, colour LCD display. 9. Should have a wireless remote control. 10. Linear footswitch control of aspiration flow rate or vacuum limit in I/A and vitrectomy mode. 11. Linear footswitch control of Ultrasound power in different US mode. 12. Should have magnetic flowability. 13. Should have preferably a #Visco# mode to remove viscoelastic from the eye. 14. Pump system # digital pulse pump. 15. Should have the facility to use vacuum level of minimum 500 mm of Hg and aspiration flow rate of minimum 50 cc/min. 16. The machine should have occlusion mode. 17. Facility for bimanual microphaco. 18. Bipolar coagulation capability. 19. The machine should be upgradable to any future technology. 20. 3 (Three) year post warranty AMC & CMC (separately) to be incorporated. 	1	NO

Tender No. & Date : KID7439P08/09

05.12.2007

- Special Notes** :
1. The bidder must submit technical literature / catalogue of the offered product in duplicate along with the offer, failing which offer may be liable for rejection.
 2. Materials must be guaranteed for fabrication, workmanship and performance for a period of 18 months from the date of despatch or 12 months from the date of commissioning whichever is earlier and relevant guarantee certificates have to be provided along with the bill / despatch documents.
 3. The equipment has to be installed and commissioned by successful bidder at OIL, Duliajan, Assam on free of cost basis.