

Bid Corrigendum

GEM/2021/B/1368749-C2

Following terms and conditions supersede all existing "Buyer added Bid Specific Terms and conditions" given in the bid document or any previous corrigendum. Prospective bidders are advised to bid as per following Terms and Conditions:

Buyer Added Bid Specific Additional Terms and Conditions

1. **Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.
2. Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address OIL INDIA LIMITED Secondary Tank Farm (STF), Madhuban Duliajan, Dibrugarh Assam India - 786602.
3. Bidder shall submit the following documents along with their bid for Vendor Code Creation:
 - a. Copy of PAN Card.
 - b. Copy of GSTIN.
 - c. Copy of Cancelled Cheque.
 - d. Copy of EFT Mandate duly certified by Bank.
4. Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorised Reseller.
5. **OPTION CLAUSE:** The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.
6. Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.
7. The buyer organization is an institution eligible for concessional rates of GST as notified by the Government of India. The goods for which bids have been invited fall under classification of GST concession and the conditions for eligibility of concession are met by the institution. A certificate to this effect will be issued by Buyer to the Seller after award of the Contract. Sellers are requested to submit their bids after accounting for the Concessional rate of GST.
Applicable Concessional rate of GST : 5%
Notification No.and date : 3/2017 dated 28/06/2017
8. **Upload Manufacturer authorization:** Wherever Authorised Distributors are submitting the bid, Manufacturers Authorisation Form (MAF)/Certificate with OEM details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid.
9. Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:
 - i) The Seller fails to comply with any material term of the Contract.
 - ii) The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.
 - iii) The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.
 - iv) The Seller becomes bankrupt or goes into liquidation.
 - v) The Seller makes a general assignment for the benefit of creditors.
 - vi) A receiver is appointed for any substantial property owned by the Seller.

vii) The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

10. While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.
11. Whereever Essentiality Certificate is applicable (PEL/ML), successful bidder should provide Proforma Invoice for processeing for EC application and material should be dispatche after receiving of EC rom DGH. In view of the same, an ATC may be incorporated in GeM, viz, "BIDDER/OEM must provide Proforma Invoice for processeing for EC application within 95 days from date of issue of GeM Contract and material should be dispatche after receiving of EC rom DGH."
12. Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
13. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
14. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for percentage of 100% of total value.
15. Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
16. Warranty period of the supplied products shall be 1 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.
17. Buyer uploaded ATC document [Click here to view the file](#).

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization. Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specification and / or terms and conditions governing the bid. Any clause incorporated by the Buyer such as demanding Tender Sample, incorporating any clause against the MSME policy and Preference to make in India Policy, mandating any Brand names or Foreign Certification, changing the default time period for Acceptance of material or payment timeline governed by OM of Department of Expenditure shall be null and void and would not be considered part of bid. Further any reference of conditions published on any external site or reference to external documents / clauses shall also be null and void.

*This document shall overwrite all previous versions of Bid Specific Additional Terms and Conditions.

[This Bid is also governed by the General Terms and Conditions](#)

Addendum No.01 dated 31.07.2021.

This document has been published to amend some of the existing technical specifications and display the minutes of pre-bid conference:

1) Amendments in existing technical specification:

| Clause ref. | Point No. | Sub Point | Existing Specification | Amended to read as |
|-------------|-----------|-----------|------------------------|---|
| Annexure-I | 4.17.2 | 5.10 | Display: Touch Screen | Display: Touch Screen/ Non-touch screen |
| Annexure-I | 4.17.3 | 4.13 | Display: Touch Screen | Display: Touch Screen/ Non-touch screen |

Tender No. GEM/2021/B/1368749 - PRESSURE AND TEMPERATURE TEST & CALIBRATION BENCH**2) Minutes of Pre-bid meeting:**

| MINUTES OF PRE-BID CONFERENCE MEETING HELD ON 29 th JULY 2021 (VIDEO CONFERENCE) AGAINST GEM TENDER NO: GEM/2021/B/1368749 - PRESSURE AND TEMPERATURE TEST & CALIBRATION BENCH | | | | |
|---|--|--|---|--|
| Attendee's:- | | | | |
| OIL'S Representative | Shri B. Mazumdar, DGM(Inst.) | M/s. Fluke Technologies Pvt Ltd. | Shri. Kaushik Thirumalachar | |
| OIL'S Representative | Shri R. Bhagowati, DGM(Inst.) | M/s. Fluke Technologies Pvt Ltd. | Shri. Abhishek Das | |
| OIL'S Representative | Shri M. Sharma, Sr. Manager(F&A) | M/s Sonepar India Pvt Ltd. | Shri. Ajit Dubey | |
| OIL'S Representative | Shri A. L. Bora, Sr. Manager Materials (FD) | M/s Sonepar India Pvt Ltd. | Shri. Pritam Sarkar | |
| OIL'S Representative | Shri R. Morang, SE(Inst.) | | | |
| | | | | |
| During the pre-bid meeting, following queries were received from M/s. Sonepar, details of discussion is as under: | | | | |
| | | | | |
| SR.NO | POINT NO. | AS PER TENDER | PARTY'S QUERY | OIL'S REPLY |
| Party's Name :M/s. Sonepar India Pvt. Ltd. | | | | |
| 1.0 | 4.17.2 (sub point 5.10) & 4.17.3 (sub point 4.13) | Pneumatic Low pressure controller & Pneumatic pressure controller (0-100 Bar) Display: Touch Screen | Fluke has launched a new product Fluke 2271A which has got better features such as in-built HART option, mA measuring option, switch test option, dual test ports etc. It doesn't come with Touch screen. If OIL is ok removing the touch screen specification from the enquiry, we will offer Fluke 2271A. | OIL shall review the same and intimate accordingly. |
| 2.0 | 4.17.2 (sub point 4.1) | Pneumatic Low pressure controller Range 1: - 350 mbar to 350 mbar | Low Pressure Calibrator Module range availability: We wish to clarify that we will be offering -600 to 600mbar pressure module for the Low Pressure Sensor Range 1 against the requested range of -350 to 350mbar. We hereby confirm that the specification are fully met by this module. | OIL confirmed that the higher range is acceptable, provided the accuracy and other technical requirements as per tender should be met. |

Tender No. GEM/2021/B/1368749 - PRESSURE AND TEMPERATURE TEST & CALIBRATION BENCH

| SR.NO | POINT NO. | AS PER TENDER | PARTY'S QUERY | OIL'S REPLY |
|-------|---------------------------|--|---|---|
| 3.0 | 4.17.15 (sub point 4.1) | Temperature Calibration Bath 4.1 Range: 35°C to 300 °C | Temperature oil bath: Due to current oil availability in market the Range of oil available is upto 275 degC. Product range will be 300degC whereas oil range will be 275 degC. Hence, we request OIL to add the operating range upto 275 degC. | OIL confirmed that the product range shall be up to 300 °C as per the tender specification. |
| 4.0 | 4.17.16 (sub point 5.9) | Dry Block Temperature Calibrator 5.9 Insert Dimension: Ø 25 mm x 150 mm (Minimum) | Dry block calibrator: We will be offering Fluke 9144 drywell that contains insert Outer diameter of 24.4mm against 25mm asked in tender. All other specifications of Fluke 9144 is superior than the RFQ requirement. Requesting confirmation regarding this as 25mm is a non-standard dimension. | OIL confirmed that the insert outer diameter as stipulated in the tender is the minimum requirement, which has to be met. |
| 5.0 | - | - | We request OIL to clarify the Payment terms and conditions. | 80% of total cost quoted will be paid against materials delivery, 20% will be paid after successful installation & commissioning. |