OIL INDIA LIMITED

(A Government of India Enterprise) P.O. Duliajan, Pin – 786602 Dist-Dibrugarh, Assam

Amendment 2 dated 19.09.2022 to Tender no. SDG1087P23 Dtd 27.07.2022.

This amendment is issued for the following:

AA) To amend the specifications of tender as below:

| Existing Clause | Amended Clause | Remarks |
|--|--|---|
| ANNEXURE – B AA: TECHNICAL SPECIFICATIONS | ANNEXURE – B AA: TECHNICAL SPECIFICATIONS: | Quantity was not mentioned earlier in the tender, which might create |
| SPECIFICATIONS: Scope of Supply: | Scope of Supply: | confusion amongst the bidders. |
| 3.5 <u>Injection Pump:</u> <u>Dual Cylinders</u> Pressure rating: 10,000 psi Min. Flow rate: 0.001 cc/min Max. Flow rate: 30 cc/min Wetted material: Hastelloy | 3.5 Injection Pumps for continuous flow: Oty: 02 nos. (min.) Should be capable of performing both Unsteady State & Steady State experiments Pressure rating: 10,000 psi Min. Flow rate: 0.001 cc/min Max. Flow rate: 30 cc/min | Moreover, it has been clarified hereby that the pumps should be capable of performing both the desired experiments as per the specifications given in the tender. |
| With Continuous delivery rate | Wetted material: Hastelloy With Continuous delivery rate | the tender. |

- BB) To extend the Bid Closing date, Technical Bid Opening date of the tender as under:
- i) Bid Closing Date & Time: 06.10.2022 at 11:00 (IST)
- ii) Technical Bid Opening Date & Time: 06.10.2022 at 14:00 (IST).
- CC) All other terms and condition of the tender remain unchanged.

Sd/(TUHIN ROY)
CMM-FD
MATERIALS DEPARTMENT
FOR GM- MATERIALS
FOR RESIDENT CHIEF EXECUTIVE