

Tender No. : DID3288L15/L2
Tender Date : 19.04.2014
Bid Closing On : 29.05.2014 at 13:00 hrs.(IST)
Bid Opening On : 29.05.2014 at 13:00 hrs.(IST)

Tender issued to following parties only:

Slno	V_Code	Vendor Name	City/Country
1	201676	NEW MEDICAL STORES	DULIAJAN
2	202450	HOSPIMEDICA INTERNATIONAL LTD.	NEW DELHI
3	209587	KAILASH LIFESCINCES PVT LTD.	NEW DELHI
4	209588	TATVA MARKETING & SERVICES PVT LTD.	NEW DELHI
5	209589	SPECTRUM MEDICAL DEVICES	NEW DELHI
6	209842	PRATISTHA ENTERPRISE	GUWAHATI

OIL INDIA LIMITED
(A Govt. of India Enterprise)
P.O. Duliajan-786602, Assam

E-mail:material@oilindia.in, Fax No.91-374-2800533

Tender No. & Date : DID3288L15/L2 19.04.2014

Bid Security Amount : INR 0.00 OR USD 0.00
(or equivalent Amount in any currency)

Bidding Type : Single Bid (Composite Bid)

Bid Closing On : 29.05.2014 at 13:00 hrs. (IST)
Bid Opening On : 29.05.2014 at 13:00 hrs. (IST)

Performance Guarantee : Not Applicable

OIL INDIA LIMITED invites Limited tenders for items detailed below:

Item No./ Mat. Code	Material Description	Quantity	UOM
10 0C000172	<p>MUSCLE STIMULATOR -1 NO</p> <p># 4 channel electrotherapy with provision of application of two or four channels for treatment of small as well as very large muscle groups and central-peripheral system.</p> <p># facility of selecting correct parameters of each channel independently.</p> <p># facility of selecting different therapeutic currents simultaneously on different channels.</p> <p>#</p> <p>Two-pole interferential , Classical interferential, Isoplanar vector, Dipole vector, Dipole vector automatic, Russian stimulation, TENS Symmetrical biphasic pulsed current, TENS Asymmetrical biphasic pulsed current, TENS Burst, Rectangular Pulsed Current, Triangular pulsed current (IG30, IG50, IG100, IG150, IG30 alternating, IG50 alternating, IG100 alternating, IG150 alternating), Microcurrent, High voltage current, Diadynamic current (MF, DF, CP, CPid, LP), Medium-freq. interrupted direct current, Continuous (galvanic) direct current.</p> <p># facility of up-grading the firmware in future, at the installation site, through a memory card.</p> <p># large TFT Color display monitor.</p> <p># The display monitor should be scratch resistant.</p> <p># The display should be adjustable to ensure good visibility from all angles.</p> <p># The unit should be of the functional ergonomic design to be equally convenient to use by both left-and right-handed persons.</p> <p># Continuous monitoring and feedback watchdog facility to assure that unit will take appropriate measures to assure patient and user safety in case of any discrepancy.</p> <p># Pre-programmed protocols: Not less than fifty.</p> <p># Free programmable positions: Unlimited.</p> <p># Should have the facility of modifying and storing the stored protocols.</p> <p># Should have the stimulating intensity of 0-140 mA in Symmetrical and Asymmetrical TENS currents.</p> <p># Should have the facility of drawing s/d curves with the display of the same on screen and translating the same into parameters that should</p>	1	NO

Tender No. & Date : DID3288L15/L2

19.04.2014

Item No./ Mat. Code	Material Description	Quantity	UOM
	<p>be used for stimulation. # The unit should have IEC/CE Standard.</p>		
<p><u>20</u> 0C000172</p>	<p>MICROWAVE DIATHERMY MACHINE Working Frequency : 24.50 MHz Modes : Continuous & Pulsed Output Power : 250 Watts Peak 1600 Watts Pulse repetition frequency : 200 Hz Display : LCD Program : Inbuilt protocol program for different indications. Memory : 80 customised memories can be stored with patients name and can be recalled at any time. Duty cycle : Adjustable from 20%-80% Protection class : 1 B IEC standard : IEC 601-1 Dimension : 40 -45x 37-40 x 94-100 cm Weight : 35-40 kg.</p> <p>The equipment should be supplied with large field and Conical Radiator as stand</p>	1	NO
<p><u>30</u> 0C000172</p>	<p>EN-CARDIO EXERCISE EQUIPMENT adjustable from a seated position: the load should be set by a simple push on a button. » The unit should provide clear feedback to the user: the force should be displayed digitally and the number of repetitions should be counted automatically » The large green letters should also visible for persons with visual limitations » The unit should be manual adjustable » With the chipcards, the resistance and number of repetitions should be set automatically » The unit should be simple to operate and accurately adjustable » The unit should have no unnecessary high peak forces » The unit should be silent during exercising » Minimal increments should be possible, thus accurate setting and smooth progression » The unit should have wide range of load setting: from extremely light to very heavy » The unit should be compact » The unit should be supplied with the following: - Installation kit - Compressor for sufficient pressure - Personal Computer - Software</p>	1	NO

Standard Notes: (1) VALIDITY : Your offer must be valid for 75 days from the date of bid opening. Offer with inadequate validity will be rejected.

(2) The offer should be submitted in Duplicate.

(3) Any sum of money due and payable to the contractor (including Security Deposit refundable to them) under this or any other contract may be appropriated by Oil India Limited and set-off against any claim of Oil India Limited (or such other person or persons contracting through Oil India Limited) for payment of a sum of money arising out of this contract or under any other contract made by the contractor with Oil India Limited (or such other person or persons contracting through Oil India Limited).

(4) In the event you authorize your dealer/stockist/channel partner to quote on your behalf, the dealer/stockist/channel partner while submitting bid should mention on the body of the envelope that they are submitting bid on your behalf.

In the event the dealer/stockist/channel partner do not mention the name of their OEM/principal on the body of the envelope, the offer shall be treated as unsolicited offer and will not be considered for opening.

The dealer/stockist/channel partner should take note of above while submitting bid on behalf of their OEM/principal.

(5) For order with F.O.R. Destination term, 100% payment against despatch documents will not be entertained. In this regards please refer payment terms in ANNEXURE-MM/TENDER/LP/01/06.

(6) To evaluate the inter-se ranking of the offers, Assam Entry Tax on purchase value will be loaded as per prevailing Government of Assam Guidelines as applicable on bid closing date. Bidders may check this with the appropriate authority while submitting their bids.

(7) Bidder must mention page no./nos. in every pages of their offer.

(8) Bidders should submit their bids (preferably in tabular form) explicitly mentioning compliance / non compliance to all the NIT terms and conditions of NIT.

(9) Bidder should clearly mention their name and address on the outside of the envelope containing their offer.

(10) In the event of receipt of only a single offer against the tender within B.C.date, OIL reserves the right to extend the B.C. date as deemed fit by the company. During the extended period, the bidders who have already submitted the bids on or before the original B.C date, shall not be permitted to revise their quotation.