

OIL INDIA LIMITED

(A Govt. Of India Enterprise) Tel :033 2230 1657, 1658
 4, India Exchange Place, Fax :91 33 2230 2596
 Kolkata-700001 E-mail :oilcalmn@oilindia.in

Tender No. & Date : KID4904L15/09 18.09.2014

Bid Security Amount : INR 0.00 OR USD 0.00
 (or equivalent Amount in any currency)

Bidding Type : Single Bid (Composite Bid)

Bid Closing On : 24.10.2014 at 14:00 hrs. (IST)
 Bid Opening On : 24.10.2014 at 14:00 hrs. (IST)

Performance Guarantee : Not Applicable

OIL INDIA LIMITED invites Limited tenders for items detailed below:

Item No./ Mat. Code	Material Description	Quantity	UOM
10 0C000172	Data logging Area Heat Stress Monitor SPECIFICATION FOR "Area Heat Stress Monitor" : * Accuracy & ranges +/- 0.1oC, from 32 oC to 40 oC * Operating Range :Temp- 0 oC to 70 oC * Temp. reading : oC & oF. * Time stamping with clock * 10 seconds data logging intervals. * Detailed time history data. * IP54 water & dust ingress protection rating * Aluminum case. Humidity- 0% to 95% * Output: RS-232 serial printer / computer interface. * The equipments should be intrinsically safe for use in a gaseous Hydrocarbon environment & DGMS approved.	1	NO
20 0C000172	Light Meter SPECIFICATION FOR "Light Meter": * Multiple wide measuring range. * MIN/MAX readings * Auto power off to save battery. * Precision silicon photodiode sensor with long cable for easy storage. *Cosine and color corrected measurements. * The equipments should be intrinsically safe for use in a gaseous Hydrocarbon environment & DGMS approved.	1	NO
30 0C000172	Personal Air Sampler Pump SPECIFICATION FOR "Personal Air Sampler Pump".: * Sample duration readout & convenient hold feature. * Short term samples over an extended period to meet TWA requirements with fewer samples. * ensured sample integrity upon low battery, restricted flow, or excess back pressure.	1	NO

Tender No. & Date : KID4904L15/09 18.09.2014

Item No./ Mat. Code	Material Description	Quantity	UOM
	<p>* Timed shutdown. * Extended intermittent sampling up to 7 days. * The equipments should be intrinsically safe for use in a gaseous Hydrocarbon environment & DGMS approved</p> <p>PERSONAL AIR SAMPLER PUMP SHOULD MEET THE FOLLOWING ACCESARRIES ALSO:</p> <p>1) Respirable dust media(Aluminum Cyclone, 37 mm # 1 no). 2) Inhalable dust sampler 1 NO 3) Filters, 1 NO EACH (i) 37 mm (PVC, 50 PK) (ii) GFA 25 mm Filter + 4) Filter Cassette 3 NOS.</p>		
<p>40 0C000172</p>	<p>Analytical Balance</p> <p>Specification of Analytical Balance: *Motorized internal calibration. (Optional in SI # 234) *ISO GLP compliance *Multiple weighing units *Standard bi-directional RS-232 interface *Polyfunction: % weighing, counting, fill mode, GSM etc. *Backlight LCD graphical display *Capacity tracking indication.</p>	1	NO

- Special Notes** : 1. Materials are most urgently required for our hospital. Please quote your best delivery.
2. Guarantee / Warranty certificate for one year will be required along with the supply.
3. Validity of offer: 75 days from the date of tender opening. Offer with validity less than 75 days will be rejected.

Tender No. : KID4904L15/09
Tender Date : 18.09.2014
Bid Closing On : 24.10.2014 at 14:00 hrs.(IST)
Bid Opening On : 24.10.2014 at 14:00 hrs.(IST)

Tender issued to following parties only:

Slno	V_Code	Vendor Name	City/Country
1	203289	SWAN ENVIRONMENTAL PVT.LTD.	HYDERABAD
2	211451	SCIENTIFIC INTERNATIONAL	DELHI
3	211452	ENSAS INDIA	VASAI,THANE
4	211453	B L HEALTH & SAFETY PVT. LTD.	NEW DELHI
5	211454	BIOMATE INDIA	NEW DELHI
6	403960	FRESH AIR	GUWAHATI