

AMENDMENT NO. 1 TO TENDER NUMBER GEM/2024/B/5356580 DTD 04.09.2024 FOR PROCUREMENT OF HIGH PRESSURE TEST UNIT

This amendment is issued to amend the following in Technical Specification as mentioned in Annexure I of ATC instead of existing.

Item/ Clause no.	Existing clause	Revised clause
Item 10. – Clause No. – 12	One (01) number 6.35 mm (¼ inch) NPT female hydraulic outlet	One (01) number 9/16”NPT female hydraulic outlet
Item 10. – Clause No. – 13	Unit mounted chart recorder for recording the test pressure on circular chart with 24 Hrs - 90 Minutes spring drive. Design should be such that there is minimum vibration transfer from the Hydraulic Pump to the chart recorder to ensure vibration free operation of the recorder pens.	Unit mounted chart recorder for recording the test pressure on circular chart with 24 hours -96 minutes spring drive. Design should be such that there is minimum vibration transfer from the Hydraulic Pump to the chart recorder to ensure vibration free operation of the recorder pens.
Item 10. – Clause No. – 15	One (01) number 20 meters (66 feet) long, 9.525 mm OD high pressure hydraulic test hose with at least 1055 Kg/Sq.Cm. (0-15000 psi) W.P. for connecting the test unit to testing equipment. The end connections of hose should be 6.35 mm (¼ inch) NPT male for both ends.	One (01) number 20 meters (66 feet) long, 9.525 mm OD high pressure hydraulic test hose with at least 1055 Kg/Sq.Cm. (0-15000 psi) W.P. for connecting the test unit to testing equipment. The end connections of hose should be be 9/16” – 18 Type – Female & ¼” NPT male for both ends.
Item 10. – Clause No. – 19	Extra charts for recording in 24 Hrs and 90 minutes recording- 100 pieces.	Extra charts for recording 24 hours -96 minutes recording- 100 pieces.

BC date extended to 09.10.2024.

All other terms and conditions of the tender remain unchanged.