

OIL INDIA LIMITED
(A Government of India Enterprise)
P.O. Duliajan, Pin – 786602
Dist-Dibrugarh, Assam

CORRIGENDUM NO. 3 DATED 27.09.2022

To

Tender No. CDI1390P23 for ‘Hiring of Calibration Services of Measuring Instruments/Equipment at OIL Duliajan’.

This Corrigendum is issued to notify the following changes:

1. The following document has been newly uploaded in the “Amendments” folder in E-portal:
 - OIL’s response to the queries of the Pre-Bid conference held on 20.09.2022.

All others terms and conditions of the Bid Document remain unchanged. Details can be viewed at www.oil-india.com.

OIL's response to the queries of Pre-Bid Conference held on 20.09.2022 through VC against Tender No. CDI1390P23 for 'Hiring of Calibration Services of Measuring Instruments/Equipment at OIL Duliajan'.

Sl. No.	Page No.	Section	Tender Clause	Clarification Sought/Recommendations	OIL's Response
1.		Section-II, Scope of Work	3.10 The Contractor must provide copy of the valid NABL traceable calibration certificate of their Master Calibrating Equipment along with the calibration certificate of OIL's equipment, otherwise calibration shall be considered void.	Whether NABL Logo is required in Calibration certificates of the Instruments calibrated by the bidder.	May please be guided by NIT Clause Nos. 3.8 and 3.10 (All test equipment should have the traceability to the National Accreditation Board for Testing and Calibration (NABL) accredited test lab. The Contractor must provide copy of the valid NABL traceable calibration certificate of their Master Calibrating Equipment along with the calibration certificate of OIL's equipment, otherwise calibration shall be considered void.)
2.		Section-II Scope of Work	3.8 Test equipment used by the contractor should be of accuracy better than the instruments being calibrated. All test equipment should have the traceability to the National Accreditation Board for Testing and Calibration (NABL) accredited test lab as per Clause No. 7.0 of OISD STD 153.	Suggestion to consider other accreditation bodies which are equivalent to NABL in tender specifications.	May please be guided by NIT Clause Nos. 3.8 and 3.10 as the same has been incorporated as per Oil Industry Safety Directorate (OISD) standard.
3.		BEC/ BRC	2.1 EXPERIENCE:	It has been apprised by the party that they have the	If the volume of the part of work that has been completed is more than or

			<p>The bidder must have experience in successfully executing/completing at least 01 (One) 'SIMILAR WORK' of minimum value of Rs. 20,99,620.00 (Rupees Twenty Lakh Ninety-Nine Thousand Six Hundred Twenty only) under single contract during the last 07 (Seven) years reckoned from the original bid closing date in Central Govt./State Govt./Public Sector Undertaking/State Govt. Enterprise/Public Limited Company.</p>	<p>technical capability to carry out the services as per tender specification, but they can't meet the requirement of 'Single contract value of Rs. 20,99,620.00' as mentioned in Tender document No. CD11390P23, BEC Clause No. 2.1. Also, they have mentioned that they have multiple running and closed contracts with value more than Rs. 20.99 lakh and have a running contract of Rs. 30.0 lakhs.</p> <p>Hence, they requested to kindly review the experience criteria of Rs.20,99,620.00 in order to participate.</p>	<p>equal to the requisite value and condition as per BEC/BRC of the tender in their said contract, then the party may participate with that credential. Further, the party was advised to check their credentials and if further clarifications is required, the party was asked to send their queries.</p>
4.	Page 4 of 9	BEC/BRC	<p>3.0 COMMERCIAL EVALUATION CRITERIA:</p>	<p>The party mentioned that as a Central Govt. entity, they shall not be able to submit Bid Security.</p>	<p>May please be guided by 3.0 COMMERCIAL EVALUATION CRITERIA:</p> <p>Bid Security is not applicable against this tender. However, bidders shall submit along with their bid a signed "Bid Security Declaration" (PROFORMA-V) provided in the tender document.</p>

Note: All bidders were advised to send their additional queries (if any) by 21.09.2022.